

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
 CERTIFICATES FOR ELECTRICAL EQUIPMENT
 (IECEE) CB SCHEME

 SYSTEME CEI D'ACCEPTATION MUTUELLE DE
 CERTIFICATS D'ESSAIS DES EQUIPEMENTS
 ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

 Product
 Produit

Dental X-ray unit (Dental CT)

 Name and address of the applicant
 Nom et adresse du demandeur

 J. Morita Manufacturing Corporation Quality Assurance Dept.
 680 Higashihama Minami-cho, FUSHIMI-KU KYOTO 612-8533
 JAPAN

 Name and address of the manufacturer
 Nom et adresse du fabricant

 J. Morita Manufacturing Corporation Quality Assurance Dept.
 680 Higashihama Minami-cho, FUSHIMI-KU KYOTO 612-8533
 JAPAN

 Name and address of the factory
 Nom et adresse de l'usine

 J. Morita Manufacturing Corporation Quality Assurance Dept.
 680 Higashihama Minami-cho, FUSHIMI-KU KYOTO 612-8533
 JAPAN

 Note: When more than one factory, please report on page 2
 Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2^{ème} page

 Ratings and principal characteristics
 Valeurs nominales et caractéristiques principales

 Additional Information on page 2

 Rated voltage: 120/220/230/240 V;
 Rated frequency: 50/60 Hz;
 *add. ratings see page 2

 Trademark (if any)
 Marque de fabrique (si elle existe)

J. Morita

 Model / Type Ref.
 Ref. De type

 X550, type EX1 (120V),
 X550, type EX2 (220/230/240V)

 Additional information (if necessary may also be
 reported on page 2)
 Les informations complémentaires (si nécessaire,,
 peuvent être indiqués sur la 2^{ème} page

 applied part type B; class I; IPX0;
 device classification acc. 93/42 MDD; IIb; TMP

 Additional Information on page 2

PUBLICATION
EDITION

 A sample of the product was tested and found
 to be in conformity with
 Un échantillon de ce produit a été essayé et a été
 considéré conforme à la

 IEC 60601-1(ed.2);am1;am2
 * collateral and particular standards see page 2

 As shown in the Test Report Ref. No. which forms part
 of this Certificate
 Comme indiqué dans le Rapport d'essais numéro de
 référence qui constitue partie de ce Certificat

2159300-2882-0002/111303

 This CB Test Certificate is issued by the National Certification Body
 Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**
VDE Prüf- und Zertifizierungsinstitut GmbH
VDE Testing and Certification Institute
 Zertifizierungsstelle / Certification

Date: 2009-12-04

Signature:



 E. Biehl

Additional information (if necessary)
Information complémentaire (si nécessaire)

add. standards:

IEC 60601-1-3(ed.1)
 IEC 60601-1-4(ed.1);am1
 IEC 60601-1-6(ed.1)
 IEC 60601-2-7(ed.2)
 IEC 60601-2-28(ed.1)
 IEC 60601-2-32(ed.1)
 IEC 60601-2-44(ed.2)

add.ratings:

Rated current: 20.4/10.4/10/9.5A Rated output 90kV, 10mA
 Rated current: 19/10.4/10/9.5A; Rated output 80 kV, 10mA
 Rated power: max. 2.2kVA (120V), 2.3kVA (220/230/240V)
 Rated output: 90kV, 10mA
 Rated power: max. 2.0kVA (120v), 2.3kVA (220/230/240V)
 Rated output: 80kV, 10mA

Amendment of A1 for correction of type error

VDE Prüf- und Zertifizierungsinstitut GmbH
VDE Testing and Certification Institute
 Zertifizierungsstelle / Certification

Date: 2009-12-04

Signature:



E. Biehl