

### DECLARATION OF CONFORMITY

We undersigned MAGNOLIA S.r.l., with registered office addressed in Via G. Natta 6/a - 43122 Parma, and headquarters office addressed in Via B. Franklin 31/a- 43122 Parma, declare under its own responsibility that the medical devices:

| Medical device                         | Risk class | Rule |
|--|------------|------|
| ANTISCHIUMOGENO DISINFETTANTE          | Ila        | 15   |
| ECO JET 1 TISSUE                       | Ila        | 15   |
| ECO JET 1 SPRAY                        | Ila        | 15   |
| ECO JET 1 SPRAY RICARICA               | Ila        | 15   |
| PULI JET PLUS 2.0 -PULI JET GENTLE 2.0 | Ila        | 15   |
| AQUA PLUS - AQUA PLUS 2.0              | Ila        | 15   |
| PULI JET PLUS NEW - PULI JET GENTLE    | Ila        | 15   |
| FAST & STERIL 3                        | IIb        | 15   |
| FAST & STERIL 3 IMPRONTE               | IIb        | 15   |

according to the Directive 93/42/EEC and further amendments, Annex IX (enforced in Italy by Legislative Decree No. 46/1997 and further amendments), as amended by the Directive 2007/47/EC (enforced in Italy by Legislative Decree No. 37/10):

- are manufactured in accordance with the Quality System that meets the requirements set out in Annex II of the aforementioned Legislative Decree, as per Certificate no. 0425-MED-003241-01 issued on 30 April 2021, expiring on 26 May 2024, by the Notified Body ICIM S.p.a. n. 0425, Piazza Don Enrico Mapelli, 75 - 20099 Sesto San Giovanni (MI).

Place

Parma

Date

12.07.2022

Legal Representative

Ing. Ennio Cattani

